

## GOBACK CATHETER VIDEO TRANSCRIPT

### **Tips and tricks: Precautions**

- Do not use polymer-coated guidewires
- To avoid damaging the guidewire, pull back the needle before pulling back the guidewire into the GoBack
- When treating anterior tibial artery occlusions, the 2.9 French GoBack should be used
- When changing needle direction inside the plaque, it is important to pull back the needle into the Catheter before rotating the knob
- The GoBack is contraindicated for use in the coronary, cerebral and carotid arteries